



**TOTAL BILIRUBIN (Diazo-sulfanilic) REAGENT KIT
V314-16D, C314-0C, C314-0A**

CONTENTS

PRODUCT NO.

PACKAGING

TOTAL BILIRUBIN (Diazo-Sulfanilic) REAGENT KIT	V314-16D	
SULFANILIC REAGENT (R1)	V314-17	3 x 13 mL
NITRITE REAGENT (R2)	V314-18	3 x 4 mL
TOTAL BILIRUBIN (Diazo-Sulfanilic) REAGENT KIT	C314-0C	
SULFANILIC REAGENT (R1)	C314-19	2 x 25 mL
NITRITE REAGENT (R2)	C314-20	2 x 7 mL
TOTAL BILIRUBIN (Diazo-Sulfanilic) REAGENT KIT	C314-0A	
SULFANILIC REAGENT (R1)	C314-01	4 x 120 mL
NITRITE REAGENT (R2)	C314-02	4 x 25 mL

REAGENT PREPARATION

These reagents are packaged ready for use. No preparation is required.

REAGENT STORAGE AND STABILITY

Store unopened reagent at 15 - 30°C (room temperature). When stored as directed, the reagent is stable until the expiration date stated on the label.

DIRECTIONS FOR USE

The Catachem Total Bilirubin test method requires two reagents: the Sulfanilic Reagent, as R1 and the Nitrite Reagent, as R2.

After opening, these reagents are stable on board the analyzer at 2-8°C (refrigerated) for 30 days. Cap both reagents tightly when not in use.

NOT FOR USE IN UNPROFESSIONAL SETTINGS.

FOR TECHNICAL ASSISTANCE PLEASE CONTACT:

catachem@catacheminc.com

www.catacheminc.com