

CREATININE REAGENTS V320-24

CONTENTS	PRODUCT	PACKAGING
CREATININE REAGENT KIT	V320-24	
DILUENT REAGENT (R1)	V320-01	3 x 13 mL
COLOR REAGENT(R2)	V320-02	3 x 13 mL

REAGENT PREPARATION

These reagents are packaged ready for use. No preparation is required.

REAGENT STORAGE AND STABILITY

Store the unopened Diluent Reagent (R1) and unopened Color Reagent (R2) at 15-30°C (room temperature) or at 2-8°C (refrigerated). When stored as directed, both reagents are stable until the expiration date stated on the label.

DIRECTIONS FOR USE

The Catachem Creatinine test method requires two reagents: Diluent Reagent as R1 and Color Reagent as R2. After opening, these reagents are stable until the expiration date stated on the label. Opened reagents may also be stored on board an automated analyzer at 2-8°C (refrigerated) or at 8-15°C. Cap reagents when not is use. NOT FOR USE IN UNPROFESSIONAL SETTINGS.

FOR TECHNICAL ASSISTANCE PLEASE CALL:

(203) 262-0330

OR EMAIL:

catachem@catacheminc.com

www.catacheminc.com