



[IN THE USA, FOR RESEARCH USE ONLY AND NOT FOR USE IN DIAGNOSTIC PROCEDURES]

**ETHANOL REAGENT KIT**  
(including Standard)  
**C701-0A;C701-0B**

<u>PRODUCT NO.</u>	<u>CONTENTS</u>	<u>PACKAGING</u>
<b>ETHANOL REAGENT KIT</b>	<b>C701-0A</b>	
ETHANOL REAGENT (R1)	C701-01	2 x 13 mL
ETHANOL REAGENT (R2)	C701-02	2 x 2.6 mL
<b>ETHANOL STANDARD 2.4% VOLUME/VOLUME</b>	C701-21	2 x 1 mL
<b>ETHANOL REAGENT KIT</b>	<b>C701-0B</b>	
ETHANOL REAGENT (R1)	C701-03	2 x 26 mL
ETHANOL REAGENT (R2)	C701-04	2 x 5.2 mL
<b>ETHANOL STANDARD 2.4% VOLUME/VOLUME</b>	C701-22	2 x 1.5 mL

**Note:** Assigned values above denote the actual values multiplied by a dilution factor of 51 as used in the Catachem ethanol assay. This dilution factor compensates for the sample dilution of 1:51 when the test is run. ( As an example the actual ethanol concentration v/v of this Standard of “2.4%” is actually 2.4/51 or 0.047%)

**REAGENT PREPARATION**

This reagent is packaged ready for use. No preparation is required.

**REAGENT STORAGE AND STABILITY**

Store reagent at 2-8°C (refrigerated). When stored as directed, the reagent is stable until the expiration date stated on the label.

**DIRECTIONS FOR USE**

The Catachem, Inc. Ethanol Reagent, after opening is stable for 30 days at 2-8°C (refrigerated). Cap the reagent when not in use. NOT FOR USE IN UNPROFESSIONAL SETTINGS.

**FOR TECHNICAL ASSISTANCE PLEASE CALL:**

**(203) 262-0330**

**OR EMAIL**

**[catachem@catacheminc.com](mailto:catachem@catacheminc.com)**