



MAGNESIUM REAGENT
C355-01

Contents	Product No.	Package
MAGNESIUM REAGENT	C355-01	1 x 125 mL

REAGENT PREPARATION

This reagent is packaged ready for use.
No preparation is required.

REAGENT STORAGE AND STABILITY

Store unopened reagent at 2-30°C.
When stored as directed, both reagents are stable until the expiration date stated on the label.

NOT FOR USE IN UNPROFESSIONAL SETTINGS

FOR TECHNICAL ASSISTANCE:
Email: catachem@catacheminc.com
Contact Form: www.catacheminc.com
Call: +1 203-262-0330



MAGNESIUM REAGENT C355-01 MANUAL/AUTOMATED APPLICATION

Intended Use

For **IN VITRO quantitative** determination of Magnesium in serum using manual or automated procedures.

Clinical Significance

Measurements of Magnesium are primarily used for diagnosing Magnesium deficiency, tetany, acute pancreatitis, hypothyroidism, chronic glomerulonephritis, aldosteronism and digitalis intoxication, dehydration, severe diabetic acidosis, uremia, as well as for monitoring the causes and treatments. ⁽¹⁾

Method History

In 1969, Chauhan and Sarkar described a procedure for the determination of traces of Magnesium in biological fluids using calmagite as an indicator. Ratge et.al. in 1986 described the use of Xylidyl Blue as a superior chromogen indicator. ⁽²⁾ This Catachem Magnesium method is based upon an inverse xylidyl blue reaction where the decrease in absorbance is read either at 630 or 650nm. The choice of these wavelengths improves precision and minimizes interferences from lipemic bilirubin and hemolyzed specimens.

Method Principle

The serum sample is mixed with the Magnesium reagent where the Magnesium ions react with the xylidyl blue indicator to form a Magnesium-chelate complex with concomitant reduction of the reagent absorbance which has a broad absorbance peak between 600 and 660nm. The decrease in absorbance is directly proportional to the Magnesium concentration in the sample. The reaction scheme illustrates the reaction that occurs in this method.

Xylidyl Blue + Mg⁺⁺ -----> Mg⁺⁺ xylidyl blue chelate

Reagent Content

The concentrations of the active ingredients in the reagent are approximately as follows:

Magnesium Reagent

Each liter contains:

Buffer	pH approx. 11.2
Xylidyl Blue	0.14 mmol
EGTA	0.10 mmol
Surfactant	

Precaution

Avoid contact of reagent with skin and eyes. Should contact occur, wash affected area with plenty of cold water. **DO NOT PIPETTE REAGENTS BY MOUTH.**

Preparation Of Working Reagent

Catachem Magnesium reagent is packaged in a ready-to-use form. No preparation is required.

Reagent Storage And Stability

Store Catachem Magnesium reagent at 2-30°C. When stored as directed, the reagent is stable until the expiration date stated on the label. Catachem Magnesium is air sensitive and should be capped when not in use. The Catachem Magnesium Reagents have been tested to reflect shipping conditions and is stable for the lifespan of the

product if frozen up to 5 times or reaching temperatures up to 40°C for up to one week.

Specimen Collection And Preparation

Test specimens should be fresh, clear, unhemolyzed sera. Serum samples stored for periods longer than eight hours should be refrigerated at 2-8°C. Under these storage conditions, samples are stable for up to 3 days.

Quality Control

To ensure optimal performance of these reagents and this procedure, we recommend systematic calibration using Catachem's Catalac (C1200-10). Assay performance should be monitored by running normal/abnormal controls concomitantly with samples. Catachem has optimized this assay using Catatrol Level I (C1200-11) and Catatrol Level II (C1200-12) and recommends their use for daily QC.

Interfering Substances

Several substances have been reported to interfere with the Magnesium method. Environmental contamination of assays will produce erroneous results. Care must be taken to use clean glassware and Magnesium-free distilled or deionized water. Anticoagulants such as oxalates and fluorides will depress Magnesium values. A summary of interferences from drugs on clinical laboratory procedures may be found by consulting D.S. Young, et al ⁽³⁾.

Expected Values

The analytical measuring range of this assay as performed below is 1.6 mg/dL to 2.6 mg/dL (0.65 to 1.05 mmol/L) when using human samples. These values serve as suggested reference points only. For veterinary samples, ranges will be dependent on the species under test. It is recommended that each laboratory establish the normal ranges for the species under study and for the geographic area in which the laboratory is located. ⁽¹⁾

Procedure

Important: Read entire procedure instructions before proceeding with assay.

Materials Required (Not Provided)

Spectrophotometer	
Match cuvettes	1 cm light path
Timer	to time incubation time
Pipette	2.0 ml for reagent
Pipette	0.02 ml for sample

Materials Provided

Catachem Magnesium reagent

Calibration

Catachem Calibrator, "Catalac" is recommended for use in the Catachem Magnesium assay. The Calibrator and the unknown should be treated in the same way while performing the Magnesium procedure.

Analytical Parameters

Wavelength	630 or 660nm
Temperature	37°C



MAGNESIUM REAGENT

V355-12, C355-01

MANUAL/AUTOMATED APPLICATION

Pathlength	1 cm
Reaction Mode	Endpoint
Reaction Time	5.0 minutes
Reagent Volume	3.0 ml
Sample Volume	0.02 ml
Total Volume	3.020 ml
Sample to Rgt Ratio	1:151

Assay Procedures

1. Pipette 3.0ml of Magnesium Reagent into each of three cuvettes marked "Calibrator", "Sample", and "Blank".
2. Pipette 0.020 ml of Calibrator or Sample into their respective cuvettes. Use 0.020ml of distilled water for the Blank. Mix all cuvettes well.
3. Incubate all cuvettes for 5.0 minutes at 37°C.
4. Set spectrophotometer wavelength at 630 or 660 nm.
5. Read the "Calibrator", "Sample", and "Blank" absorbances.
6. Calculate the Magnesium concentration (mg/dL) in the Sample(s), as shown in calculations and results below.

Calculations And Results

$$\text{Mg}^{++} \text{ (mg/dL)} = \frac{\text{Ba} - \text{Sa}}{\text{Ba} - \text{Ca}} \times \text{Calibrator (mg/dL)}$$

Where: Ba = Blank absorbance
 Sa = Sample absorbance
 Ca = Calibrator

ASSAY OD

Example: Sample		1.46
Calibrator		1.45
Blank		1.70
Calibrator	=	4.0 mg/dL

$$\text{Calcium (mg/dL)} = \frac{0.240}{0.250} \times 4.0 \text{ mg/dL} = 3.84 \text{ mg/dL}$$

NOTE: Samples with Magnesium concentrations greater than 6.2 mg/dL should be diluted with physiological saline and reassayed. Results should be adjusted for dilution.

Method Performance Characteristics

Sensitivity: 0.05-0.07 absorbance units/mg/dL

Linear Range: 0.0 - 6.2 mg/dL (2.5 mmol/L)

Precision: Within-run and day-to-day precision is summarized below:

Precision Study

Mg ⁺⁺	TOTAL		WITHIN-RUN	
	MEAN	SD	SD	CV
mg/dL	mg/dL	%	mg/dL	%
1.7	0.13	7.70	0.12	7.06
3.1	0.17	5.48	0.14	4.52
6.2	0.20	3.22	0.16	2.58

Correlation

A comparison of this method using an automated analyzer resulted in the following regression statistics:

Range	=	0.64 – 4.46 mg/dL
N	=	40
Y	=	0.956 x -0.05
r	=	0.994

References

1. Fundamentals of Clinical Chemistry. Edited by Norbert Teitz. WB Saunders, Philadelphia (1976).
2. Ratge, D., Kohse, K. P. et Wisser, H., Measurement of Magnesium in Serum and Urine with a Random Access Analyzer by Use of a Modified Xylidyl Blue-1 Procedure, Clin. Chem. Acta, 159 (1986) 197-203.
3. Young D.S, Pestaner LD, Gibberman V. Clin. Chem. 21, 5 (1975)